



# Daily Calibration Log for Automatic Leak Tester (ZUTR10003)

Automatic Leak Tester is compatible with ACMI, Fujifilm, Fuse, Olympus, Pentax, Richard Wolf & Storz endoscopes.

### Calibrate the AUTOMATIC LEAK TESTER daily:

1. Connect the Calibration Gauge (ZUTR10003TG) to the Automatic Leak Tester.
2. Turn the Automatic Leak Tester on.
3. This unit is calibrated to produce air pressure between 120–196 mmHg
4. If the gauge shows air pressure output between 120-196 mmHg, then it passes the daily calibration test. If it is not within this range, please contact Zutron Medical.

Month/Year: \_\_\_\_\_

Serial #: \_\_\_\_\_

DAY	PASS	FAIL	INITIAL	COMMENTS
1	PASS	FAIL		
2	PASS	FAIL		
3	PASS	FAIL		
4	PASS	FAIL		
5	PASS	FAIL		
6	PASS	FAIL		
7	PASS	FAIL		
8	PASS	FAIL		
9	PASS	FAIL		
10	PASS	FAIL		
11	PASS	FAIL		
12	PASS	FAIL		
13	PASS	FAIL		
14	PASS	FAIL		
15	PASS	FAIL		
16	PASS	FAIL		
17	PASS	FAIL		
18	PASS	FAIL		
19	PASS	FAIL		
20	PASS	FAIL		
21	PASS	FAIL		
22	PASS	FAIL		
23	PASS	FAIL		
24	PASS	FAIL		
25	PASS	FAIL		
26	PASS	FAIL		
27	PASS	FAIL		
28	PASS	FAIL		
29	PASS	FAIL		
30	PASS	FAIL		
31	PASS	FAIL		

Please refer to the Automatic Leak Tester IFU for more details. Or contact your local distributor or Zutron Medical for further assistance: 877-343-5873